

PATENT

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Applicant(s)	Ben Z. Cohen, et al.	Examiner:	Kenneth Bomberg
Serial No.:	10/069,682	Group Art Unit:	3754
Confirmation No.:	3571	Docket:	1181-3 PCT/US/RCE
Filed:	August 7, 2002	Dated:	October 25, 2004
For:	MICRODISPENSING PUMP		

Mail Stop RCE
Commissioner for Patents
P.O. Box 1450
Alexandria, Virginia 22313-1450

*I hereby certify this correspondence is being deposited
with the United States Postal Service as first class mail,
postpaid in an envelope, addressed to: Mail Stop RCE
Commissioner for Patents, P.O. Box 1450,
Alexandria, Virginia 22313-1450
On: October 25, 2004*

Signed: Mary J. MacDonald
Mary J. MacDonald

SUPPLEMENTAL AMENDMENT

Sir:

A final Office Action was issued in the above-identified application on August 4, 2004. An Amendment After Final Rejection was filed in response on October 4, 2004. In an Advisory Action mailed on October 18, 2004, the Examiner denied entry of the October 4 Amendment and stated that Applicant has failed to point out where the limitation of "the diaphragm remaining in the deflected position" finds support in the corresponding specification. Filed co-currently herewith is a Request for Continued Examination (RCE), which requests entry of the previously denied amendment. In response to the Examiner's comment seeking support for the aforementioned limitation, the Examiner's attention is drawn to page 15, lines 8-11 of Applicants' specification where it states in part,

Application No.: 10/069,682
Filed August 7, 2004
Page 2


Instead of priming the dispenser by pressing the cap several times, the user presses the diaphragm 90, which deflects inwards into the fluid reservoir 28 **and remains in that position.**

(Emphasis applied).

Applicants respectfully submit that this passage provides adequate disclosure for the proposed amendment.

Favorable action is earnestly solicited. If there are any questions of if additional information is required, the Examiner is respectfully requested to contact Applicants' attorney at the number listed below.

Respectfully submitted,


Ludomir A. Budzyn
Registration No.: 40,540
Attorney for Applicant(s)

HOFFMANN & BARON, LLP
6900 Jericho Turnpike
Syosset, New York 11791
(973) 331-1700